
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R750
SN 500919I00001
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 000246
7:15 21/07/2020
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:54
PRESSURIZING 3:00
STERILIZING 4:18
136.0 C 220kPa 4:18
136.0 C 221kPa 4:48
136.4 C 222kPa 5:18
136.4 C 222kPa 5:48
136.2 C 218kPa 6:18
136.4 C 223kPa 6:48
136.4 C 224kPa 7:18
136.3 C 223kPa 7:48
136.3 C 223kPa 7:48
VENTING 7:48
AIR DRYING 8:43
CYCLE COMPLETE 8:47

Digital Signature #
A04BB9C0291C50CF
