
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
1.2uS / 0.7ppm
CYCLE NUMBER 000605
13:15 26/08/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:38
PRESSURIZING 2:46
STERILIZING 3:55
136.4 C 231kPa 3:55
136.2 C 229kPa 4:25
136.4 C 232kPa 4:55
136.2 C 231kPa 5:25
136.3 C 230kPa 5:55
136.4 C 232kPa 6:25
136.3 C 232kPa 6:55
136.2 C 232kPa 7:25
136.2 C 232kPa 7:25
VENTING 7:25
AIR DRYING 8:15
CYCLE COMPLETE 22:07

Digital Signature #
59E65F2E7DB95A2A
