

-----  
Midmark M11 - v1.0.5  
Total Cycles: 4000

Name: RITTER2  
Sterilizer ID: V2313228  
10/04/2024 10:14 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 154.5 F 0.0  
2:00 151.6 F 1.2  
4:00 154.4 F 4.0  
6:00 178.7 F 7.5  
8:00 195.9 F 11.3  
10:00 212.8 F 4.5  
12:00 223.5 F 8.7  
14:00 234.5 F 13.0  
16:00 233.8 F 9.6  
18:00 237.5 F 10.2  
20:00 244.1 F 13.1  
22:00 255.0 F 18.9  
24:00 264.9 F 25.0  
26:00 271.6 F 29.4

STERILIZING  
mm:ss Degrees PSI  
0:00 271.2 F 29.3  
1:00 271.6 F 29.8  
2:00 272.2 F 30.3  
3:00 272.6 F 30.3  
4:00 272.9 F 30.7

Min 271.0 F 29.2  
Max 272.9 F 30.7

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 207.7 F 0.0  
5:00 198.5 F 0.0  
10:00 227.8 F 0.0  
15:00 226.0 F 0.0  
20:00 222.7 F 0.0  
25:00 231.0 F 0.0  
30:00 228.5 F 0.0  
Drying Complete

FILLING: 0:60  
HEATING: 26:10  
STERILIZING: 4:00  
VENTING: 3:52  
DRYING: 30:00  
TOTAL CYCLE: 01:05:02  
CYCLE COMPLETE

10/04/2024 11:18 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_