
Midmark M11 - v1.0.5
Total Cycles: 2462

Name: RITTER2
Sterilizer ID: V2313228
06/05/2023 09:32 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 107.0 F 0.0
2:00 110.1 F 0.0
4:00 116.4 F 0.1
6:00 153.5 F 2.5
8:00 173.3 F 6.4
10:00 210.3 F 10.6
12:00 216.5 F 5.7
14:00 223.3 F 9.1
16:00 235.8 F 11.7
18:00 237.8 F 10.9
20:00 235.4 F 8.9
22:00 248.6 F 15.2
24:00 259.9 F 21.4
26:00 270.2 F 28.1

STERILIZING
mm:ss Degrees PSI
0:00 271.8 F 29.0
1:00 271.5 F 29.4
2:00 272.6 F 29.6
3:00 271.6 F 29.0
4:00 271.4 F 29.4

Min 271.1 F 28.8
Max 272.7 F 29.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 196.5 F 0.0
5:00 213.5 F 0.0
10:00 226.8 F 0.0
15:00 214.3 F 0.0
20:00 212.5 F 0.0
25:00 213.3 F 0.0
30:00 218.9 F 0.0
Drying Complete

FILLING: 0:49
HEATING: 26:47
STERILIZING: 4:00
VENTING: 3:10
DRYING: 30:00
TOTAL CYCLE: 01:04:46
CYCLE COMPLETE

06/05/2023 10:36 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____