
Midmark M11 - v1.0.0
Total Cycles: 5470

Name: RITTER 1
Sterilizer ID: V2178198
04/02/2025 01:44 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 135.6 F 0.0
2:00 138.6 F 2.0
4:00 146.3 F 4.8
6:00 175.3 F 7.8
8:00 190.1 F 11.7
10:00 214.1 F 3.0
12:00 214.7 F 6.8
14:00 224.4 F 10.5
16:00 235.7 F 8.5
18:00 239.4 F 10.6
20:00 237.6 F 8.6
22:00 246.3 F 13.2
24:00 256.2 F 18.4
26:00 264.8 F 23.9
28:00 272.0 F 28.6

STERILIZING
mm:ss Degrees PSI
0:00 271.7 F 28.6
1:00 271.9 F 28.6
2:00 271.9 F 28.8
3:00 272.1 F 29.0
4:00 271.7 F 28.9

Min 271.6 F 28.5
Max 272.2 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 190.6 F 0.0
5:00 196.5 F 0.0
10:00 202.1 F 0.0
15:00 211.9 F 0.0
20:00 220.3 F 0.0
25:00 211.9 F 0.0
30:00 213.4 F 0.0
Drying Complete

FILLING: 0:52
HEATING: 28:07
STERILIZING: 4:00
VENTING: 4:00
DRYING: 30:00
TOTAL CYCLE: 01:06:59
CYCLE COMPLETE

04/02/2025 02:50 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____