

-----  
Midmark M11 - v1.0.0  
Total Cycles: 792

Name: RITTER 1  
Sterilizer ID: V2178198  
04/21/2021 11:07 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 145.5 F 0.0  
2:00 148.3 F 1.9  
4:00 165.9 F 5.3  
6:00 193.6 F 10.1  
8:00 224.7 F 4.7  
10:00 236.8 F 10.3  
12:00 232.9 F 7.2  
14:00 245.2 F 13.1  
16:00 242.1 F 11.3  
18:00 255.6 F 18.2  
20:00 267.4 F 25.4  
22:00 271.8 F 28.7

STERILIZING  
mm:ss Degrees PSI  
0:00 271.7 F 28.8  
1:00 271.9 F 28.7  
2:00 271.9 F 28.7  
3:00 272.4 F 28.8  
4:00 272.6 F 29.0

Min 271.6 F 28.6  
Max 272.8 F 29.3

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 177.8 F 0.0  
5:00 237.4 F 0.0  
10:00 237.3 F 0.0  
15:00 227.8 F 0.0  
Dry Cycle Aborted

FILLING: 1:00  
HEATING: 22:02  
STERILIZING: 4:00  
VENTING: 2:54  
DRYING: 16:22  
TOTAL CYCLE: 00:46:18

04/21/2021 11:53 AM  
ACCEPTED: X  
REJECTED:  
Operator: U01