
Brampton, ON
Canada

STATIM 5000 S501R713
SN 511019D00004
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 001398
8:06 09/08/2021
UNWRAPPED
134 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:42
PRESSURIZING 5:24
STERILIZING 6:45
136.0 C 220kPa 6:45
136.4 C 219kPa 7:15
136.3 C 222kPa 7:45
136.1 C 221kPa 8:15
136.4 C 223kPa 8:45
136.4 C 223kPa 9:15
136.4 C 223kPa 9:45
136.5 C 217kPa 10:15
VENTING 10:16
AIR DRYING 11:16
CYCLE COMPLETE 11:22

Digital Signature #
BD63E5F79D9A5056
