
Midmark M11 - v1.0.5
Total Cycles: 539

Name: RITTER2
Sterilizer ID: V2313228
11/11/2021 07:27 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 74.2 F 0.0
2:00 74.3 F 0.0
4:00 77.6 F 0.0
6:00 115.1 F 0.2
8:00 146.8 F 1.0
10:00 174.1 F 4.3
12:00 198.3 F 7.7
14:00 206.2 F 12.1
16:00 223.5 F 6.3
18:00 235.4 F 10.9
20:00 234.3 F 7.6
22:00 244.2 F 13.5
24:00 243.5 F 12.4
26:00 255.7 F 19.0
28:00 266.9 F 25.8

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 29.0
1:00 272.7 F 29.4
2:00 271.5 F 29.4
3:00 272.3 F 29.3
4:00 271.5 F 29.5

Min 271.1 F 28.8
Max 273.0 F 29.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 177.8 F 0.0
5:00 217.8 F 0.0
10:00 200.4 F 0.0
15:00 189.7 F 0.0
20:00 219.8 F 0.0
25:00 228.4 F 0.0
30:00 216.8 F 0.0
Drying Complete

FILLING: 0:56
HEATING: 29:30
STERILIZING: 4:00
VENTING: 3:04
DRYING: 30:00
TOTAL CYCLE: 01:07:30
CYCLE COMPLETE

11/11/2021 08:34 AM
ACCEPTED: X
REJECTED:
Operator: U02