
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.1uS / 0.0ppm
CYCLE NUMBER 001916
8:32 07/08/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:00
PRESSURIZING 2:07
STERILIZING 3:16
136.0 C 222kPa 3:16
136.3 C 224kPa 3:46
136.4 C 226kPa 4:16
136.4 C 223kPa 4:46
136.2 C 224kPa 5:16
136.2 C 224kPa 5:46
136.3 C 222kPa 6:16
136.4 C 225kPa 6:46
136.4 C 225kPa 6:46
VENTING 6:46
AIR DRYING 7:37
CYCLE COMPLETE 22:42

Digital Signature #
A4D733001F7B14ED
