
Midmark M11 - v1.0.0
Total Cycles: 1673

Name: RITTER 1
Sterilizer ID: V2178198
02/24/2022 05:04 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 161.7 F 0.0
2:00 157.8 F 1.8
4:00 177.9 F 4.6
6:00 199.6 F 8.3
8:00 213.5 F 13.3
10:00 227.5 F 7.2
12:00 239.9 F 12.0
14:00 237.2 F 9.2
16:00 239.5 F 9.5
18:00 248.9 F 14.3
20:00 260.9 F 21.0
22:00 271.7 F 28.2

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.1
1:00 272.2 F 28.2
2:00 272.3 F 28.7
3:00 271.4 F 28.2
4:00 271.8 F 28.4

Min 271.3 F 28.0
Max 272.7 F 28.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 193.3 F 0.0
5:00 207.7 F 0.0
10:00 203.4 F 0.0
15:00 208.7 F 0.0
20:00 213.6 F 0.0
25:00 210.3 F 0.0
30:00 219.1 F 0.0
Drying Complete

FILLING: 1:16
HEATING: 22:30
STERILIZING: 4:00
VENTING: 3:15
DRYING: 30:00
TOTAL CYCLE: 01:01:01
CYCLE COMPLETE

02/24/2022 06:04 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____