
Midmark M11 - v1.0.0
Total Cycles: 2542

Name: RITTER 1
Sterilizer ID: V2178198
11/10/2022 02:43 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 193.4 F 0.0
2:00 196.4 F 4.1
4:00 208.2 F 9.0
6:00 223.2 F 8.0
8:00 230.1 F 9.2
10:00 238.9 F 9.3
12:00 242.7 F 12.4
14:00 241.5 F 11.1
16:00 254.8 F 17.9
18:00 266.4 F 25.0

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.4
1:00 272.3 F 28.8
2:00 271.9 F 28.8
3:00 272.3 F 28.7
4:00 271.9 F 28.8

Min 271.5 F 28.3
Max 272.7 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 213.8 F 0.0
5:00 227.3 F 0.0
10:00 235.0 F 0.0
15:00 227.9 F 0.0
20:00 227.8 F 0.0
25:00 229.0 F 0.0
30:00 227.7 F 0.0
Drying Complete

FILLING: 1:44
HEATING: 19:31
STERILIZING: 4:00
VENTING: 3:15
DRYING: 30:00
TOTAL CYCLE: 00:58:30
CYCLE COMPLETE

11/10/2022 03:41 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____