

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.2uS / 0.1ppm  
CYCLE NUMBER 000650  
8:25 13/09/2021  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 3:10  
PRESSURIZING 3:13  
STERILIZING 4:24  
136.4 C 231kPa 4:24  
136.4 C 233kPa 4:54  
136.4 C 229kPa 5:24  
136.3 C 232kPa 5:54  
136.3 C 231kPa 6:24  
136.3 C 232kPa 6:54  
136.4 C 232kPa 7:24  
136.3 C 230kPa 7:54  
136.3 C 230kPa 7:54  
VENTING 7:54  
AIR DRYING 8:49  
CYCLE COMPLETE 8:51  
-----

Digital Signature #  
71227D489F5A8684  
-----