
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 002477
11:42 30/01/2025
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:45
PRESSURIZING 2:51
STERILIZING 3:55
136.4 C 225kPa 3:55
136.2 C 223kPa 4:25
136.4 C 225kPa 4:55
136.4 C 224kPa 5:25
136.1 C 222kPa 5:55
136.2 C 223kPa 6:25
136.3 C 224kPa 6:55
136.3 C 224kPa 7:25
136.3 C 224kPa 7:25
VENTING 7:26
AIR DRYING 8:17
CYCLE COMPLETE 30:15

Digital Signature #
CB1D04B81BB0E3BE
