
Midmark M11 - v1.0.5
Total Cycles: 601

Name: RITTER2
Sterilizer ID: V2313228
12/06/2021 09:01 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 71.7 F 0.0
2:00 90.6 F 0.0
4:00 142.0 F 0.0
6:00 183.4 F 1.1
8:00 207.9 F 4.1
10:00 215.2 F 7.6
12:00 221.1 F 11.5
14:00 221.2 F 4.8
16:00 217.6 F 9.9
18:00 231.3 F 9.4
20:00 235.0 F 10.6
22:00 234.6 F 8.1
24:00 246.2 F 14.4
26:00 258.0 F 20.7
28:00 267.8 F 27.3

STERILIZING
mm:ss Degrees PSI
0:00 272.1 F 30.6
1:00 272.0 F 30.6
2:00 271.9 F 30.8
3:00 271.9 F 30.6
4:00 271.7 F 30.6

Min 271.7 F 30.5
Max 272.6 F 30.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 187.6 F 0.0
5:00 236.2 F 0.0
10:00 222.6 F 0.0
15:00 224.7 F 0.0
20:00 175.6 F 0.0
25:00 173.5 F 0.0
30:00 175.1 F 0.0
Drying Complete

FILLING: 1:04
HEATING: 29:54
STERILIZING: 4:00
VENTING: 3:01
DRYING: 30:00
TOTAL CYCLE: 01:08:00
CYCLE COMPLETE

12/06/2021 10:45 AM
ACCEPTED: X
REJECTED:
Operator: U02