
Midmark M11 - v1.0.5
Total Cycles: 213

Name: RITTER2
Sterilizer ID: V2313228
07/15/2021 09:16 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 166.2 F 0.0
2:00 162.4 F 2.8
4:00 178.5 F 6.4
6:00 194.2 F 11.0
8:00 217.9 F 5.2
10:00 225.4 F 10.2
12:00 231.3 F 8.3
14:00 242.4 F 12.9
16:00 242.4 F 11.8
18:00 254.7 F 18.3
20:00 265.8 F 25.0

STERILIZING
mm:ss Degrees PSI
0:00 271.0 F 28.9
1:00 273.0 F 29.7
2:00 271.9 F 29.7
3:00 272.1 F 29.3
4:00 272.5 F 29.9

Min 271.0 F 28.7
Max 273.2 F 30.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 179.2 F 0.0
5:00 236.1 F 0.0
Dry Cycle Aborted

FILLING: 1:08
HEATING: 21:39
STERILIZING: 4:00
VENTING: 3:10
DRYING: 6:42
TOTAL CYCLE: 00:36:39

07/15/2021 09:52 AM
ACCEPTED: X
REJECTED:
Operator: U02