
Midmark M11 - v1.0.0
Total Cycles: 720

Name: MIDMARK
Sterilizer ID: V2178198
03/02/2021 03:09 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 176.4 F 0.0
2:00 171.4 F 3.5
4:00 186.6 F 8.2
6:00 207.1 F 8.7
8:00 224.0 F 9.0
10:00 233.9 F 8.8
12:00 243.1 F 12.5
14:00 241.9 F 11.2
16:00 256.2 F 18.5
18:00 268.5 F 26.2

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.7
1:00 271.8 F 28.6
2:00 272.0 F 28.6
3:00 272.3 F 28.7
4:00 272.5 F 28.9

Min 271.5 F 28.5
Max 272.9 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 206.7 F 0.0
5:00 230.7 F 0.0
10:00 223.2 F 0.0
15:00 220.1 F 0.0
20:00 197.9 F 0.0
25:00 204.4 F 0.0
30:00 204.8 F 0.0
Drying Complete

FILLING: 1:27
HEATING: 19:44
STERILIZING: 4:00
VENTING: 2:57
DRYING: 30:00
TOTAL CYCLE: 00:58:08
CYCLE COMPLETE

03/02/2021 04:07 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____