

-----  
Midmark M11 - v1.0.0  
Total Cycles: 3338

Name: RITTER 1  
Sterilizer ID: V2178198  
06/05/2023 03:43 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 210.9 F 0.0  
2:00 206.7 F 2.1  
4:00 227.1 F 8.1  
6:00 236.8 F 11.6  
8:00 236.7 F 10.4  
10:00 237.4 F 8.1  
12:00 240.7 F 10.2  
14:00 249.0 F 14.6  
16:00 261.6 F 21.7  
18:00 272.8 F 28.8

STERILIZING  
mm:ss Degrees PSI  
0:00 271.6 F 28.1  
1:00 272.3 F 28.5  
2:00 272.3 F 28.7  
3:00 272.0 F 28.3  
4:00 272.4 F 28.9  
  
Min 271.6 F 28.0  
Max 272.7 F 28.9

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 212.5 F 0.0  
5:00 224.9 F 0.0  
10:00 231.0 F 0.0  
15:00 228.5 F 0.0  
20:00 226.2 F 0.0  
25:00 224.5 F 0.0  
30:00 222.3 F 0.0  
Drying Complete

FILLING: 2:25  
HEATING: 18:20  
STERILIZING: 4:00  
VENTING: 3:17  
DRYING: 30:00  
TOTAL CYCLE: 00:58:02  
CYCLE COMPLETE

06/05/2023 04:40 PM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_