
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R750
SN 500919I00001
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 000800
7:26 12/11/2020
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:50
PRESSURIZING 2:54
STERILIZING 3:55
136.4 C 222kPa 3:55
136.4 C 221kPa 4:25
136.1 C 222kPa 4:55
136.4 C 221kPa 5:25
136.1 C 221kPa 5:55
136.2 C 222kPa 6:25
136.4 C 222kPa 6:55
136.4 C 219kPa 7:25
136.4 C 219kPa 7:25
VENTING 7:26
AIR DRYING 8:23
CYCLE COMPLETE 8:39

Digital Signature #
B68BB72968F321CA
