
Midmark M11 - v1.0.0
Total Cycles: 4227

Name: RITTER 1
Sterilizer ID: V2178198
01/24/2024 05:10 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 168.6 F 0.0
2:00 170.5 F 3.3
4:00 187.7 F 8.1
6:00 214.3 F 11.1
8:00 226.9 F 8.4
10:00 236.6 F 10.4
12:00 240.8 F 11.4
14:00 238.5 F 9.6
16:00 252.6 F 16.5
18:00 264.5 F 23.4

STERILIZING
mm:ss Degrees PSI
0:00 271.1 F 28.1
1:00 272.8 F 28.8
2:00 271.4 F 28.0
3:00 272.4 F 28.9
4:00 272.5 F 28.6

Min 271.1 F 27.9
Max 272.9 F 29.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 196.7 F 0.0
5:00 200.4 F 0.0
10:00 195.2 F 0.0
15:00 193.6 F 0.0
20:00 191.0 F 0.0
25:00 187.2 F 0.0
30:00 183.4 F 0.0
Drying Complete

FILLING: 1:40
HEATING: 19:56
STERILIZING: 4:00
VENTING: 3:13
DRYING: 30:00
TOTAL CYCLE: 00:58:49
CYCLE COMPLETE

01/24/2024 06:08 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____