
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.1uS / 0.0ppm
CYCLE NUMBER 002428
10:26 17/01/2025
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 1:51
PRESSURIZING 1:58
STERILIZING 3:02
136.5 C 224kPa 3:02
136.2 C 224kPa 3:32
136.3 C 225kPa 4:02
136.2 C 224kPa 4:32
136.4 C 225kPa 5:02
136.3 C 225kPa 5:32
136.3 C 225kPa 6:02
136.3 C 225kPa 6:32
136.3 C 225kPa 6:32
VENTING 6:32
AIR DRYING 7:24
CYCLE COMPLETE 14:35

Digital Signature #
DB6079A3B01B0D7E
