
Midmark M11 - v1.0.5
Total Cycles: 3473

Name: RITTER2
Sterilizer ID: V2313228
03/25/2024 09:01 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 68.4 F 0.0
2:00 68.9 F 0.0
4:00 73.6 F 0.0
6:00 103.7 F 0.8
8:00 138.7 F 3.1
10:00 170.5 F 6.3
12:00 186.2 F 10.4
14:00 202.3 F 6.4
16:00 222.0 F 6.3
18:00 227.3 F 10.2
20:00 234.7 F 9.0
22:00 239.9 F 11.2
24:00 236.1 F 8.8
26:00 249.6 F 15.2
28:00 260.6 F 21.5
30:00 270.9 F 28.3

STERILIZING
mm:ss Degrees PSI
0:00 271.1 F 28.6
1:00 272.3 F 29.0
2:00 271.9 F 29.2
3:00 271.7 F 28.6
4:00 272.4 F 29.4

Min 271.1 F 28.5
Max 272.7 F 29.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 177.8 F 0.0
5:00 231.1 F 0.0
10:00 224.5 F 0.0
15:00 189.2 F 0.0
20:00 172.2 F 0.0
25:00 171.7 F 0.0
30:00 169.8 F 0.0
Drying Complete

FILLING: 0:51
HEATING: 30:42
STERILIZING: 4:00
VENTING: 3:15
DRYING: 30:00
TOTAL CYCLE: 01:08:48
CYCLE COMPLETE

03/25/2024 10:09 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____