
Midmark M11 - v1.0.0
Total Cycles: 5136

Name: RITTER 1
Sterilizer ID: V2178198
11/27/2024 03:12 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 84.3 F 0.0
2:00 84.8 F 0.0
4:00 105.2 F 0.0
6:00 137.0 F 0.2
8:00 161.1 F 0.3
10:00 186.1 F 0.6
12:00 206.1 F 3.0
14:00 218.2 F 6.6
16:00 227.7 F 10.2
18:00 235.4 F 9.0
20:00 238.3 F 10.9
22:00 235.2 F 8.4
24:00 242.2 F 10.9
26:00 246.6 F 13.4
28:00 257.9 F 19.4
30:00 268.0 F 25.8

STERILIZING
mm:ss Degrees PSI
0:00 271.7 F 28.3
1:00 272.3 F 28.9
2:00 272.7 F 29.0
3:00 272.2 F 28.6
4:00 271.7 F 28.3

Min 271.6 F 28.2
Max 272.8 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 192.7 F 0.0
5:00 203.1 F 0.0
10:00 198.1 F 0.0
15:00 199.6 F 0.0
20:00 225.9 F 0.0
25:00 214.7 F 0.0
30:00 222.2 F 0.0
Drying Complete

FILLING: 0:44
HEATING: 31:32
STERILIZING: 4:00
VENTING: 3:48
DRYING: 30:00
TOTAL CYCLE: 01:10:05
CYCLE COMPLETE

11/27/2024 04:22 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____