
Midmark M11 - v1.0.5
Total Cycles: 1196

Name: RITTER2
Sterilizer ID: V2313228
06/16/2022 11:29 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 179.1 F 0.0
2:00 171.4 F 2.7
4:00 194.8 F 7.5
6:00 214.7 F 13.3
8:00 230.0 F 8.5
10:00 238.9 F 11.4
12:00 240.6 F 11.8
14:00 239.9 F 10.7
16:00 253.4 F 17.5
18:00 265.0 F 24.5

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.9
1:00 272.9 F 29.7
2:00 271.9 F 28.9
3:00 271.9 F 29.5
4:00 272.7 F 29.6

Min 271.1 F 28.7
Max 272.9 F 29.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 191.0 F 0.0
5:00 229.8 F 0.0
10:00 224.8 F 0.0
15:00 218.3 F 0.0
20:00 179.6 F 0.0
25:00 177.7 F 0.0
30:00 180.4 F 0.0

Drying Complete

FILLING: 2:22
HEATING: 19:52
STERILIZING: 4:00
VENTING: 3:08
DRYING: 30:00
TOTAL CYCLE: 00:59:22
CYCLE COMPLETE

06/16/2022 12:28 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____