
Midmark M11 - v1.0.5
Total Cycles: 619

Name: RITTER2
Sterilizer ID: V2313228
12/09/2021 10:39 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 189.1 F 0.0
2:00 185.1 F 3.2
4:00 197.9 F 8.4
6:00 220.3 F 8.9
8:00 229.6 F 9.8
10:00 234.1 F 8.7
12:00 243.4 F 12.1
14:00 245.8 F 13.5
16:00 259.0 F 20.6
18:00 270.4 F 28.2

STERILIZING
mm:ss Degrees PSI
0:00 271.8 F 29.0
1:00 272.0 F 28.9
2:00 272.4 F 29.2
3:00 272.8 F 29.6
4:00 271.9 F 29.2

Min 271.7 F 28.8
Max 272.9 F 29.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 189.3 F 0.0
5:00 231.0 F 0.0
10:00 226.4 F 0.0
15:00 227.2 F 0.0
20:00 227.8 F 0.0
25:00 224.8 F 0.0
30:00 222.6 F 0.0
Drying Complete

FILLING: 2:29
HEATING: 18:43
STERILIZING: 4:00
VENTING: 2:60
DRYING: 30:00
TOTAL CYCLE: 00:58:12
CYCLE COMPLETE

12/09/2021 11:38 AM
ACCEPTED: X
REJECTED:
Operator: U02