
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 001376
10:06 05/12/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 4:16
PRESSURIZING 4:17
STERILIZING 5:33
136.0 C 220kPa 5:33
136.3 C 223kPa 6:03
136.4 C 224kPa 6:33
136.1 C 222kPa 7:03
136.4 C 225kPa 7:33
136.3 C 223kPa 8:03
136.3 C 223kPa 8:33
136.2 C 220kPa 9:03
136.2 C 220kPa 9:03
VENTING 9:03
AIR DRYING 10:01
CYCLE COMPLETE 21:12

Digital Signature #
09DFD1BA1B8B83FB
