
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 002183
12:18 29/10/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:07
PRESSURIZING 2:11
STERILIZING 3:11
136.4 C 223kPa 3:11
136.4 C 226kPa 3:41
136.3 C 225kPa 4:11
136.4 C 226kPa 4:41
136.2 C 223kPa 5:11
136.4 C 226kPa 5:41
136.4 C 227kPa 6:11
136.3 C 225kPa 6:41
136.3 C 225kPa 6:41
VENTING 6:41
AIR DRYING 7:32
CYCLE COMPLETE 37:32

Digital Signature #
54001173B9DB75C6
