

-----  
Midmark M11 - v1.0.0  
Total Cycles: 3878

Name: RITTER 1  
Sterilizer ID: V2178198  
10/10/2023 08:10 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 68.7 F 0.0  
2:00 93.0 F 0.0  
4:00 77.6 F 0.0  
6:00 108.4 F 0.1  
8:00 140.2 F 0.3  
10:00 165.7 F 3.1  
12:00 203.0 F 6.1  
14:00 212.2 F 9.7  
16:00 221.2 F 13.3  
18:00 223.5 F 7.0  
20:00 233.6 F 11.0  
22:00 231.8 F 7.1  
24:00 243.5 F 12.6  
26:00 240.4 F 10.8  
28:00 253.1 F 17.0  
30:00 263.8 F 23.2  
32:00 271.8 F 28.3

STERILIZING  
mm:ss Degrees PSI  
0:00 271.3 F 28.4  
1:00 271.9 F 28.7  
2:00 272.4 F 29.1  
3:00 271.4 F 28.3  
4:00 272.1 F 28.9  
  
Min 271.3 F 28.3  
Max 272.4 F 29.1

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 191.5 F 0.0  
5:00 224.2 F 0.0  
10:00 241.8 F 0.0  
15:00 235.0 F 0.0  
20:00 227.6 F 0.0  
25:00 228.7 F 0.0  
30:00 218.2 F 0.0  
  
Drying Complete

FILLING: 0:49  
HEATING: 32:07  
STERILIZING: 4:00  
VENTING: 3:24  
DRYING: 30:00  
TOTAL CYCLE: 01:10:20  
CYCLE COMPLETE

10/10/2023 09:20 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_