

-----  
Midmark M11 - v1.0.0  
Total Cycles: 3832

Name: RITTER 1  
Sterilizer ID: V2178198  
09/28/2023 08:55 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 185.3 F 0.0  
2:00 182.0 F 1.9  
4:00 191.3 F 5.9  
6:00 209.1 F 11.4  
8:00 224.0 F 6.6  
10:00 236.1 F 11.9  
12:00 235.9 F 9.1  
14:00 239.0 F 9.7  
16:00 247.7 F 13.9  
18:00 259.9 F 20.6  
20:00 270.9 F 27.7

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 28.1  
1:00 272.3 F 28.5  
2:00 272.1 F 28.8  
3:00 271.9 F 28.1  
4:00 272.6 F 28.9  
  
Min 271.4 F 28.0  
Max 272.7 F 28.9

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 208.5 F 0.0  
5:00 196.9 F 0.0  
10:00 201.2 F 0.0  
15:00 202.2 F 0.0  
20:00 233.9 F 0.0  
25:00 228.3 F 0.0  
30:00 226.7 F 0.0  
Drying Complete

FILLING: 1:29  
HEATING: 20:42  
STERILIZING: 4:00  
VENTING: 3:19  
DRYING: 30:00  
TOTAL CYCLE: 00:59:30  
CYCLE COMPLETE

09/28/2023 09:54 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_