

-----  
Midmark M11 - v1.0.0  
Total Cycles: 5407

Name: RITTER 1  
Sterilizer ID: V2178198  
03/11/2025 10:57 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 135.9 F 0.0  
2:00 135.8 F 1.7  
4:00 145.3 F 4.6  
6:00 168.6 F 8.3  
8:00 191.8 F 12.6  
10:00 210.7 F 3.9  
12:00 221.4 F 7.9  
14:00 232.7 F 12.1  
16:00 233.2 F 8.2  
18:00 244.4 F 13.4  
20:00 242.4 F 11.6  
22:00 253.9 F 17.5  
24:00 264.3 F 23.8  
26:00 271.3 F 28.6

STERILIZING  
mm:ss Degrees PSI  
0:00 271.6 F 29.0  
1:00 272.4 F 29.4  
2:00 271.5 F 29.1  
3:00 272.1 F 29.4  
4:00 271.3 F 29.0

Min 271.3 F 28.9  
Max 272.8 F 29.6

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 192.7 F 0.0  
5:00 207.5 F 0.0  
10:00 237.0 F 0.0  
15:00 235.8 F 0.0  
20:00 229.8 F 0.0  
25:00 229.9 F 0.0  
30:00 221.6 F 0.0  
Drying Complete

FILLING: 0:51  
HEATING: 26:55  
STERILIZING: 4:00  
VENTING: 3:53  
DRYING: 30:00  
TOTAL CYCLE: 01:05:39  
CYCLE COMPLETE

03/11/2025 12:02 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_