
Midmark M11 - v1.0.5
Total Cycles: 4100

Name: RITTER2
Sterilizer ID: V2313228
11/07/2024 06:15 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 163.4 F 0.0
2:00 165.2 F 2.9
4:00 191.3 F 7.7
6:00 220.0 F 13.0
8:00 223.1 F 8.2
10:00 235.8 F 12.4
12:00 241.3 F 11.6
14:00 239.3 F 10.5
16:00 253.3 F 17.3
18:00 264.6 F 24.2
20:00 271.7 F 29.6

STERILIZING
mm:ss Degrees PSI
0:00 271.9 F 29.5
1:00 272.2 F 29.4
2:00 272.9 F 29.6
3:00 272.8 F 29.9
4:00 272.4 F 30.1

Min 271.8 F 29.4
Max 273.3 F 30.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 182.3 F 0.0
5:00 231.7 F 0.0
10:00 205.5 F 0.0
15:00 181.7 F 0.0
20:00 175.6 F 0.0
25:00 172.1 F 0.0
30:00 169.1 F 0.0
Drying Complete

FILLING: 1:52
HEATING: 20:29
STERILIZING: 4:00
VENTING: 3:34
DRYING: 30:00
TOTAL CYCLE: 00:59:55
CYCLE COMPLETE

11/07/2024 07:14 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____