
Midmark M11 - v1.0.0
Total Cycles: 3455

Name: RITTER 1
Sterilizer ID: V2178198
06/28/2023 07:33 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 73.1 F 0.0
2:00 82.2 F 0.0
4:00 91.1 F 0.0
6:00 127.8 F 0.6
8:00 158.3 F 2.4
10:00 181.5 F 4.6
12:00 192.6 F 7.8
14:00 207.2 F 11.4
16:00 215.3 F 4.1
18:00 226.1 F 7.7
20:00 236.5 F 11.8
22:00 234.0 F 8.3
24:00 244.4 F 12.2
26:00 244.1 F 12.3
28:00 256.2 F 18.5
30:00 266.9 F 25.0

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.1
1:00 272.6 F 28.7
2:00 271.5 F 28.1
3:00 272.8 F 28.8
4:00 271.5 F 28.0

Min 271.4 F 28.0
Max 272.8 F 28.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 166.6 F 0.0
5:00 202.6 F 0.0
10:00 205.4 F 0.0
15:00 198.1 F 0.0
20:00 198.3 F 0.0
25:00 198.9 F 0.0
30:00 198.8 F 0.0
Drying Complete

FILLING: 0:45
HEATING: 31:32
STERILIZING: 4:00
VENTING: 3:10
DRYING: 30:00
TOTAL CYCLE: 01:09:28
CYCLE COMPLETE

06/28/2023 08:42 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____