
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 001400
12:04 25/03/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:41
PRESSURIZING 2:48
STERILIZING 3:53
136.0 C 220kPa 3:53
136.4 C 222kPa 4:23
136.2 C 223kPa 4:53
136.2 C 223kPa 5:23
136.3 C 224kPa 5:53
136.2 C 223kPa 6:23
136.4 C 224kPa 6:53
136.4 C 222kPa 7:23
136.4 C 222kPa 7:23
VENTING 7:23
AIR DRYING 8:16
CYCLE COMPLETE 8:23

Digital Signature #
57915C77E407954D
