
Midmark M11 - v1.0.0
Total Cycles: 1636

Name: RITTER 1
Sterilizer ID: V2178198
02/15/2022 12:40 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 180.9 F 0.0
2:00 181.7 F 2.7
4:00 193.0 F 6.8
6:00 224.7 F 11.8
8:00 228.1 F 7.3
10:00 235.7 F 12.8
12:00 239.2 F 10.3
14:00 234.9 F 8.1
16:00 250.7 F 15.3
18:00 262.9 F 22.2
20:00 272.3 F 28.2

STERILIZING
mm:ss Degrees PSI
0:00 271.8 F 28.4
1:00 272.4 F 28.7
2:00 272.6 F 29.0
3:00 271.7 F 28.6
4:00 271.9 F 28.5

Min 271.4 F 28.2
Max 272.9 F 29.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 199.2 F 0.0
5:00 239.0 F 0.0
10:00 230.0 F 0.0
15:00 226.8 F 0.0
20:00 226.2 F 0.0
25:00 228.8 F 0.0
30:00 232.3 F 0.0
Drying Complete

FILLING: 1:48
HEATING: 20:41
STERILIZING: 4:00
VENTING: 3:03
DRYING: 30:00
TOTAL CYCLE: 00:59:32
CYCLE COMPLETE

02/15/2022 01:44 PM
ACCEPTED: X
REJECTED:
Operator: U02