
Midmark M11 - v1.0.5
Total Cycles: 706

Name: RITTER2
Sterilizer ID: V2313228
01/19/2022 01:13 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 194.4 F 0.0
2:00 190.1 F 4.3
4:00 204.7 F 10.0
6:00 221.4 F 6.3
8:00 234.7 F 12.3
10:00 237.8 F 10.7
12:00 238.1 F 10.1
14:00 252.9 F 17.4
16:00 265.2 F 24.8

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 29.1
1:00 272.3 F 29.5
2:00 272.0 F 29.8
3:00 272.0 F 29.2
4:00 272.2 F 29.9

Min 271.4 F 29.0
Max 272.7 F 30.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 197.0 F 0.0
5:00 223.3 F 0.0
10:00 245.9 F 0.0
15:00 234.3 F 0.0
20:00 225.3 F 0.0
25:00 225.4 F 0.0
30:00 224.2 F 0.0

Drying Complete

FILLING: 2:02
HEATING: 17:41
STERILIZING: 4:00
VENTING: 3:08
DRYING: 30:00
TOTAL CYCLE: 00:56:51
CYCLE COMPLETE

01/19/2022 02:09 PM
ACCEPTED:
REJECTED:

Reason: _____
Operator: _____