

-----  
Midmark M11 - v1.0.5  
Total Cycles: 1596

Name: RITTER2  
Sterilizer ID: V2313228  
10/14/2022 07:26 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 74.0 F 0.0  
2:00 77.6 F 0.0  
4:00 113.4 F 0.0  
6:00 129.6 F 0.1  
8:00 148.9 F 0.2  
10:00 174.2 F 1.0  
12:00 202.8 F 4.2  
14:00 210.7 F 7.9  
16:00 222.0 F 11.5  
18:00 224.8 F 6.2  
20:00 235.8 F 10.9  
22:00 235.5 F 7.9  
24:00 243.7 F 12.4  
26:00 243.9 F 12.7  
28:00 255.6 F 18.9  
30:00 266.1 F 25.4

STERILIZING  
mm:ss Degrees PSI  
0:00 271.4 F 29.1  
1:00 272.5 F 29.7  
2:00 271.6 F 29.0  
3:00 272.5 F 29.7  
4:00 271.7 F 28.9

Min 271.4 F 28.9  
Max 272.9 F 29.9

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 206.4 F 0.0  
5:00 208.6 F 0.0  
10:00 212.8 F 0.0  
15:00 213.2 F 0.0  
20:00 212.7 F 0.0  
25:00 213.0 F 0.0  
30:00 212.6 F 0.0  
Drying Complete

FILLING: 0:51  
HEATING: 31:43  
STERILIZING: 4:00  
VENTING: 3:07  
DRYING: 30:00  
TOTAL CYCLE: 01:09:41  
CYCLE COMPLETE

10/14/2022 08:35 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_