

-----  
Midmark M11 - v1.0.0  
Total Cycles: 2035

Name: RITTER 1  
Sterilizer ID: V2178198  
06/20/2022 10:38 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 161.4 F 0.0  
2:00 162.0 F 2.3  
4:00 171.9 F 5.4  
6:00 189.6 F 9.6  
8:00 212.7 F 7.4  
10:00 229.6 F 8.5  
12:00 240.4 F 12.8  
14:00 238.2 F 10.2  
16:00 235.0 F 8.0  
18:00 248.2 F 14.3  
20:00 259.8 F 20.8  
22:00 270.7 F 27.7

STERILIZING  
mm:ss Degrees PSI  
0:00 271.4 F 28.3  
1:00 272.3 F 28.9  
2:00 271.5 F 28.4  
3:00 272.2 F 28.8  
4:00 271.4 F 28.3

Min 271.4 F 28.2  
Max 272.4 F 29.1

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 193.3 F 0.0  
5:00 213.7 F 0.0  
10:00 233.8 F 0.0  
15:00 224.9 F 0.0  
20:00 223.9 F 0.0  
25:00 235.8 F 0.0  
30:00 236.4 F 0.0  
Drying Complete

FILLING: 0:53  
HEATING: 22:48  
STERILIZING: 4:00  
VENTING: 3:09  
DRYING: 30:00  
TOTAL CYCLE: 01:00:50  
CYCLE COMPLETE

06/20/2022 11:38 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_