
Midmark M11 - v1.0.5
Total Cycles: 2744

Name: RITTER2
Sterilizer ID: V2313228
08/09/2023 12:53 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 204.3 F 0.0
2:00 201.1 F 4.0
4:00 208.9 F 10.0
6:00 222.2 F 6.4
8:00 235.3 F 11.6
10:00 235.7 F 9.6
12:00 237.0 F 9.1
14:00 250.0 F 15.9
16:00 262.2 F 23.0
18:00 272.7 F 29.6

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 29.1
1:00 272.6 F 29.7
2:00 271.6 F 29.2
3:00 272.7 F 29.7
4:00 271.7 F 29.0

Min 271.5 F 28.9
Max 272.7 F 30.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 204.4 F 0.0
5:00 213.9 F 0.0
10:00 224.0 F 0.0
15:00 228.9 F 0.0
20:00 228.2 F 0.0
25:00 228.3 F 0.0
30:00 229.8 F 0.0
Drying Complete

FILLING: 2:19
HEATING: 18:17
STERILIZING: 4:00
VENTING: 3:12
DRYING: 30:00
TOTAL CYCLE: 00:57:48
CYCLE COMPLETE

08/09/2023 01:50 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____