
Midmark M11 - v1.0.0
Total Cycles: 2129

Name: RITTER 1
Sterilizer ID: V2178198
07/14/2022 12:10 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 165.7 F 0.0
2:00 166.1 F 2.9
4:00 186.8 F 7.6
6:00 210.4 F 13.4
8:00 223.0 F 8.1
10:00 236.9 F 13.3
12:00 238.0 F 10.8
14:00 237.2 F 8.9
16:00 250.8 F 15.8
18:00 262.6 F 22.6
20:00 272.4 F 28.9

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.5
1:00 272.2 F 28.7
2:00 272.9 F 29.3
3:00 271.7 F 28.4
4:00 272.7 F 29.1

Min 271.6 F 28.4
Max 273.0 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 209.3 F 0.0
5:00 213.7 F 0.0
10:00 223.5 F 0.0
15:00 227.1 F 0.0
20:00 229.4 F 0.0
25:00 195.8 F 0.0
30:00 190.2 F 0.0
Drying Complete

FILLING: 2:02
HEATING: 20:13
STERILIZING: 4:00
VENTING: 3:12
DRYING: 30:00
TOTAL CYCLE: 00:59:27
CYCLE COMPLETE

07/14/2022 01:09 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____