
Midmark M11 - v1.0.5
Total Cycles: 116

Name: RITTER2
Sterilizer ID: V2313228
06/18/2021 08:18 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 74.6 F 0.0
2:00 74.9 F 0.2
4:00 78.3 F 1.0
6:00 97.4 F 2.8
8:00 122.3 F 5.3
10:00 150.4 F 8.7
12:00 171.7 F 12.5
14:00 198.4 F 4.5
16:00 208.6 F 8.1
18:00 221.0 F 12.0
20:00 228.1 F 8.1
22:00 238.9 F 13.1
24:00 239.7 F 11.5
26:00 251.4 F 17.3
28:00 261.5 F 23.3
30:00 271.4 F 29.7

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 29.8
1:00 272.3 F 30.7
2:00 272.7 F 30.3
3:00 271.8 F 29.7
4:00 271.8 F 30.3

Min 271.5 F 29.7
Max 272.8 F 30.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 188.7 F 0.0
5:00 196.4 F 0.0
10:00 207.2 F 0.0
15:00 211.2 F 0.0
20:00 211.6 F 0.0
25:00 213.2 F 0.0
30:00 212.9 F 0.0
Drying Complete

FILLING: 0:55
HEATING: 30:45
STERILIZING: 4:00
VENTING: 3:15
DRYING: 30:00
TOTAL CYCLE: 01:08:55
CYCLE COMPLETE

06/18/2021 09:26 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____