

-----  
Midmark M11 - v1.0.5  
Total Cycles: 1705

Name: RITTER2  
Sterilizer ID: V2313228  
11/10/2022 08:00 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 77.3 F 0.0  
2:00 79.7 F 0.0  
4:00 90.5 F 0.0  
6:00 111.8 F 0.0  
8:00 165.9 F 0.4  
10:00 193.6 F 1.1  
12:00 207.3 F 3.8  
14:00 219.3 F 7.0  
16:00 230.9 F 10.5  
18:00 236.2 F 8.6  
20:00 237.6 F 10.9  
22:00 237.0 F 8.5  
24:00 241.3 F 10.7  
26:00 246.7 F 13.8  
28:00 258.4 F 20.2  
30:00 268.9 F 26.9

STERILIZING  
mm:ss Degrees PSI  
0:00 271.1 F 28.8  
1:00 272.4 F 29.2  
2:00 271.4 F 29.0  
3:00 272.3 F 29.2  
4:00 271.3 F 28.9  
  
Min 271.1 F 28.6  
Max 272.6 F 29.6

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 207.9 F 0.0  
5:00 205.5 F 0.0  
10:00 215.8 F 0.0  
15:00 224.7 F 0.0  
20:00 228.4 F 0.0  
25:00 225.4 F 0.0  
30:00 222.2 F 0.0  
Drying Complete

FILLING: 0:50  
HEATING: 31:09  
STERILIZING: 4:00  
VENTING: 3:08  
DRYING: 30:00  
TOTAL CYCLE: 01:09:07  
CYCLE COMPLETE

11/10/2022 09:09 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_