
Midmark M11 - v1.0.0
Total Cycles: 3838

Name: RITTER 1
Sterilizer ID: V2178198
09/29/2023 08:05 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 69.9 F 0.0
2:00 82.2 F 0.0
4:00 89.3 F 0.0
6:00 113.5 F 0.0
8:00 142.5 F 0.6
10:00 170.8 F 2.7
12:00 206.8 F 5.2
14:00 213.3 F 8.8
16:00 221.2 F 12.2
18:00 224.0 F 5.7
20:00 235.2 F 10.0
22:00 237.5 F 9.3
24:00 242.3 F 11.6
26:00 238.8 F 9.6
28:00 252.1 F 16.0
30:00 263.1 F 22.4
32:00 272.5 F 28.4

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.0
1:00 272.5 F 28.4
2:00 271.9 F 28.4
3:00 272.1 F 28.1
4:00 272.3 F 28.7

Min 271.2 F 27.8
Max 272.7 F 28.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 182.2 F 0.0
5:00 221.9 F 0.0
10:00 240.5 F 0.0
15:00 228.9 F 0.0
20:00 228.8 F 0.0
25:00 228.6 F 0.0
30:00 223.6 F 0.0

Drying Complete

FILLING: 0:46
HEATING: 32:17
STERILIZING: 4:00
VENTING: 3:16
DRYING: 30:00
TOTAL CYCLE: 01:10:19
CYCLE COMPLETE

09/29/2023 09:15 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____