
Midmark M11 - v1.0.5
Total Cycles: 1713

Name: RITTER2
Sterilizer ID: V2313228
11/10/2022 06:43 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 174.9 F 0.0
2:00 176.6 F 0.3
4:00 183.8 F 3.2
6:00 205.1 F 8.3
8:00 217.8 F 11.4
10:00 235.2 F 10.5
12:00 234.6 F 8.6
14:00 239.5 F 10.0
16:00 248.8 F 15.3
18:00 261.7 F 22.6
20:00 272.9 F 29.9

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 29.2
1:00 273.0 F 30.0
2:00 271.9 F 29.3
3:00 271.7 F 29.7
4:00 272.8 F 29.9

Min 271.2 F 29.1
Max 273.0 F 30.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 185.5 F 0.0
5:00 216.3 F 0.0
10:00 245.4 F 0.0
15:00 244.6 F 0.0
20:00 237.2 F 0.0
25:00 226.7 F 0.0
30:00 224.8 F 0.0
Drying Complete

FILLING: 1:13
HEATING: 20:21
STERILIZING: 4:00
VENTING: 3:16
DRYING: 30:00
TOTAL CYCLE: 00:58:50
CYCLE COMPLETE

11/10/2022 07:41 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____