
Midmark M11 - v1.0.5
Total Cycles: 1817

Name: RITTER2
Sterilizer ID: V2313228
12/07/2022 12:32 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 179.5 F 0.0
2:00 173.0 F 3.1
4:00 178.2 F 6.9
6:00 200.2 F 11.9
8:00 215.6 F 6.0
10:00 230.3 F 11.2
12:00 231.9 F 8.5
14:00 240.2 F 10.7
16:00 246.0 F 13.8
18:00 258.6 F 20.8
20:00 269.9 F 28.0

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 29.2
1:00 272.7 F 29.8
2:00 271.6 F 29.0
3:00 272.4 F 30.0
4:00 272.3 F 29.5

Min 271.4 F 29.0
Max 272.7 F 30.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 200.4 F 0.0
5:00 225.2 F 0.0
10:00 229.4 F 0.0
15:00 229.3 F 0.0
20:00 227.1 F 0.0
25:00 227.1 F 0.0
30:00 226.7 F 0.0
Drying Complete

FILLING: 1:14
HEATING: 20:57
STERILIZING: 4:00
VENTING: 3:05
DRYING: 30:00
TOTAL CYCLE: 00:59:16
CYCLE COMPLETE

12/07/2022 01:31 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____