

-----  
Midmark M11 - v1.0.0  
Total Cycles: 3688

Name: RITTER 1  
Sterilizer ID: V2178198  
08/28/2023 09:36 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 183.6 F 0.0  
2:00 171.3 F 2.2  
4:00 177.3 F 5.7  
6:00 204.3 F 10.7  
8:00 220.1 F 4.9  
10:00 232.1 F 10.1  
12:00 233.5 F 8.1  
14:00 244.3 F 12.7  
16:00 241.9 F 11.1  
18:00 254.7 F 17.5  
20:00 265.8 F 24.3

STERILIZING  
mm:ss Degrees PSI  
0:00 271.9 F 28.1  
1:00 272.6 F 28.7  
2:00 271.2 F 28.2  
3:00 272.4 F 28.4  
4:00 271.5 F 28.4

Min 271.0 F 27.9  
Max 272.6 F 28.8

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 187.7 F 0.0  
5:00 233.9 F 0.0  
10:00 231.9 F 0.0  
15:00 227.6 F 0.0  
20:00 226.3 F 0.0  
25:00 228.3 F 0.0  
30:00 216.9 F 0.0  
Drying Complete

FILLING: 1:31  
HEATING: 21:34  
STERILIZING: 4:00  
VENTING: 3:19  
DRYING: 30:00  
TOTAL CYCLE: 01:00:24  
CYCLE COMPLETE

08/28/2023 10:36 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_