
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 002189
8:41 09/08/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:48
PRESSURIZING 3:49
STERILIZING 4:58
136.4 C 222kPa 4:58
136.1 C 222kPa 5:28
136.2 C 221kPa 5:58
136.2 C 220kPa 6:28
136.3 C 220kPa 6:58
136.4 C 222kPa 7:28
136.2 C 222kPa 7:58
136.4 C 222kPa 8:28
136.4 C 222kPa 8:28
VENTING 8:28
AIR DRYING 9:24
CYCLE COMPLETE 9:25

Digital Signature #
3AE381D23B2DE61F
