
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 001161
13:27 08/07/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:42
PRESSURIZING 2:48
STERILIZING 3:56
136.4 C 223kPa 3:56
136.4 C 226kPa 4:26
136.4 C 225kPa 4:56
136.1 C 224kPa 5:26
136.5 C 226kPa 5:56
136.4 C 225kPa 6:26
136.4 C 226kPa 6:56
136.4 C 226kPa 7:26
136.4 C 226kPa 7:26
VENTING 7:26
AIR DRYING 8:16
CYCLE COMPLETE 23:16

Digital Signature #
6EAD18FC3FC801DC
