
Midmark M11 - v1.0.0
Total Cycles: 2174

Name: RITTER 1
Sterilizer ID: V2178198
07/26/2022 06:51 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 75.0 F 0.0
2:00 75.5 F 0.0
4:00 83.4 F 0.0
6:00 115.2 F 0.7
8:00 146.5 F 3.2
10:00 173.8 F 6.7
12:00 193.3 F 10.4
14:00 204.0 F 7.8
16:00 220.0 F 6.1
18:00 230.9 F 9.7
20:00 234.4 F 11.0
22:00 238.9 F 10.2
24:00 235.6 F 8.4
26:00 247.4 F 13.8
28:00 258.7 F 19.9
30:00 269.0 F 26.4

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.2
1:00 272.4 F 28.8
2:00 271.2 F 28.2
3:00 272.4 F 28.9
4:00 271.4 F 28.1

Min 271.2 F 28.1
Max 272.6 F 29.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 182.3 F 0.0
5:00 179.3 F 0.0
10:00 145.5 F 0.0
15:00 136.8 F 0.0
20:00 132.8 F 0.0
25:00 129.3 F 0.0
30:00 127.3 F 0.0
Drying Complete

FILLING: 0:58
HEATING: 31:11
STERILIZING: 4:00
VENTING: 3:09
DRYING: 30:00
TOTAL CYCLE: 01:09:18
CYCLE COMPLETE

07/26/2022 08:00 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____