
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R750
SN 500919I00001
Unit # : 000
WATER QUALITY
0.4uS / 0.2ppm
CYCLE NUMBER 000813
14:45 12/11/2020
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:42
PRESSURIZING 2:49
STERILIZING 3:54
136.3 C 221kPa 3:54
136.4 C 225kPa 4:24
136.1 C 220kPa 4:54
136.2 C 221kPa 5:24
136.0 C 222kPa 5:54
136.4 C 224kPa 6:24
136.4 C 223kPa 6:54
136.0 C 221kPa 7:24
136.0 C 221kPa 7:24
VENTING 7:24
AIR DRYING 8:16
CYCLE COMPLETE 13:55

Digital Signature #
5932BC7136999A94
