
Midmark M11 - v1.0.5
Total Cycles: 3134

Name: RITTER2
Sterilizer ID: V2313228
11/16/2023 01:39 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 125.1 F 0.0
2:00 129.5 F 1.0
4:00 150.2 F 3.7
6:00 175.6 F 7.7
8:00 197.3 F 12.2
10:00 212.2 F 6.7
12:00 226.2 F 11.6
14:00 233.4 F 8.3
16:00 242.1 F 11.7
18:00 244.7 F 12.8
20:00 256.6 F 19.3
22:00 267.6 F 26.1
24:00 271.5 F 29.3

STERILIZING
mm:ss Degrees PSI
0:00 271.1 F 29.4
1:00 272.3 F 29.6
2:00 272.0 F 30.0
3:00 271.1 F 29.3
4:00 272.9 F 30.6

Min 271.0 F 29.2
Max 274.0 F 30.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 171.8 F 0.0
5:00 211.5 F 0.0
10:00 191.9 F 0.0
15:00 189.4 F 0.0
20:00 185.6 F 0.0
25:00 183.4 F 0.0
30:00 181.9 F 0.0
Drying Complete

FILLING: 1:01
HEATING: 24:06
STERILIZING: 4:00
VENTING: 3:34
DRYING: 30:00
TOTAL CYCLE: 01:02:40
CYCLE COMPLETE

11/16/2023 02:41 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____