

-----  
Midmark M11 - v1.0.5  
Total Cycles: 2939

Name: RITTER2  
Sterilizer ID: V2313228  
09/27/2023 01:54 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 148.4 F 0.0  
2:00 146.5 F 0.0  
4:00 165.8 F 1.5  
6:00 193.6 F 5.9  
8:00 215.4 F 10.4  
10:00 226.8 F 6.8  
12:00 235.5 F 10.4  
14:00 234.5 F 8.2  
16:00 245.0 F 13.2  
18:00 243.3 F 12.1  
20:00 255.4 F 18.4  
22:00 266.3 F 25.0

STERILIZING  
mm:ss Degrees PSI  
0:00 271.9 F 28.6  
1:00 272.2 F 29.4  
2:00 271.9 F 28.7  
3:00 271.9 F 29.3  
4:00 272.0 F 29.0

Min 271.2 F 28.5  
Max 272.7 F 29.5

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 176.4 F 0.0  
5:00 240.4 F 0.0  
10:00 227.3 F 0.0  
15:00 228.1 F 0.0  
20:00 211.4 F 0.0  
25:00 206.1 F 0.0  
30:00 208.6 F 0.0  
Drying Complete

FILLING: 0:54  
HEATING: 23:32  
STERILIZING: 4:00  
VENTING: 3:28  
DRYING: 30:00  
TOTAL CYCLE: 01:01:55  
CYCLE COMPLETE

09/27/2023 02:55 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_